

Proton Precession Magnetometer Fluid Testing
Two second Polarization Time GELLER

<http://www.gellerlabs.com>

Sorted by Fall Time

Fluid	Amplitude		Fall Time	
	pk pk avg	pk pk std	fall t avg	fall t std
Acetone	1.90	0.12	2.16	0.11
Clorox Bleach	3.08	0.18	2.10	0.08
Distilled Water (Walgreens)	2.96	0.09	2.09	0.06
Tap Water (single filtered)	2.89	0.09	2.06	0.10
Tap Water (double filtered)	2.80	0.10	1.98	0.06
Distilled Water (homemade)	2.77	0.15	1.96	0.70
Simple Green (cleaner)	3.46	0.07	1.91	0.05
RainX Old Bottle	3.29	0.11	1.89	0.05
RainX 2in1	3.36	0.10	1.87	0.07
Wine Vinegar	3.43	0.05	1.87	0.03
RainX	3.39	0.12	1.84	0.07
Prestone Windshield Fluid	3.30	0.09	1.81	0.03
Rubber Cement	2.78	0.19	1.68	0.06
Charcoal Lighter Fluid	3.68	0.17	1.65	0.04
Alcohol 90%	2.97	0.12	1.52	0.06
Alcohol 70%	3.64	0.18	1.46	0.03

Sorted by Amplitude

Fluid	Amplitude		Fall Time	
	pk pk avg	pk pk std	fall t avg	fall t std
Charcoal Lighter Fluid	3.68	0.17	1.65	0.04
Alcohol 70%	3.64	0.18	1.46	0.03
Simple Green (cleaner)	3.46	0.07	1.91	0.05
Wine Vinegar	3.43	0.05	1.87	0.03
RainX	3.39	0.12	1.84	0.07
RainX 2in1	3.36	0.10	1.87	0.07
Prestone Windshield Fluid	3.30	0.09	1.81	0.03
RainX Old Bottle	3.29	0.11	1.89	0.05
Clorox Bleach	3.08	0.18	2.10	0.08
Alcohol 90%	2.97	0.12	1.52	0.06
Distilled Water (Walgreens)	2.96	0.09	2.09	0.06
Tap Water (single filtered)	2.89	0.09	2.06	0.10
Tap Water (double filtered)	2.80	0.10	1.98	0.06
Rubber Cement	2.78	0.19	1.68	0.06
Distilled Water (homemade)	2.77	0.15	1.96	0.70
Acetone	1.90	0.12	2.16	0.11

FALL TIME VS AMPLITUDE FOR VARIOUS PPM FLUIDS

